

# Aicar HPLC Analysis Report

Sample: **Aicar**

Buffer A: 0.1% TFA in 100% water (v/v)

Buffer B: 0.1% TFA in 80% acetonitrile + 20% water (v/v)


Gradient: 0-60% Buffer B in 20min

Flow: 1ml/min

Wavelength: 220nm

Column: Venusil C18 5um 4.6\*250mm

Chromatogram



---

Signal	Peak#	RT(min)	Area%	Area(mAU*s)	Height(mAU)
--------	-------	---------	-------	-------------	-------------

---

	1	8.006	99.91	6897526	686228
--	---	-------	-------	---------	--------

	2	8.300	0.08972	6194	1578
<hr/>					
Total			100	6903720	687806

PROPERTY OF PEPTIDES WAREHOUSE