

HPLC Analysis Report

Sample: Dapoxetine HCL

CAS: 164656-23-9

Buffer A: 0.04% TFA in 100% water(v/v)

Buffer B: 0.04% TFA in 100% acetonitrile (v/v)

Gradient: 10-80% Buffer B in 25 min then 100% Buffer B for 5 min.


Flow: 1ml/min Wavelength:220nm

Column: Alltima C18,4.6*250mm,5um

Batch No.: KB20141203

Date: December 8, 2014

Chromatogram



Peak#	RetTime(Min)	Width(Min)	Area(mAU*s)	Height(mAU)	Area(%)
1	18.965	BB	0.3456	4.66380e4	99.9193
2	22.707	BB	0.2070	37.68803	0.0807
Totals			4.66757e4	2235.44702	100

MS Spectrum

*MSD1 SPC, time=0.273:1.109 of C:\CHEM32\1\DATA\141209\140210 2014-12-09 10-21-49\SIG1000002.D ES-API, Pos, Scan, Fr

